



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE

United States Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/587,310

07/26/2006

Takashi Kawabata

06-477/LH

4527

1933

7590

03/03/2009

FRISHAUF, HOLTZ, GOODMAN & CHICK, PC

220 Fifth Avenue

16TH Floor

NEW YORK, NY 10001-7708

EXAMINER

PEFFLEY, MICHAEL F

ART UNIT

PAPER NUMBER

3739

MAIL DATE

DELIVERY MODE

03/03/2009

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/587,310

Applicant(s)

KAWABATA ET AL.

Examiner

Michael Peffley

Art Unit

3739

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 22 September 2008.
2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 9-24 is/are pending in the application.
4a) Of the above claim(s) _____ is/are withdrawn from consideration.
5) ☐ Claim(s) _____ is/are allowed.
6) ☒ Claim(s) 9-24 is/are rejected.
7) ☐ Claim(s) _____ is/are objected to.
8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
10) ☒ The drawing(s) filed on 26 July 2006 is/are: a) ☐ accepted or b) ☒ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
a) ☒ All b) ☐ Some * c) ☐ None of:
1. ☒ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
2) ☐ Notice of Draftperson's Patent Drawing Review (PTO-948)
3) ☒ Information Disclosure Statement(s) (PTO-8508)
Paper No(s)/Mail Date 7/3/08; 7/3/08; 9/25/08
4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date _____
5) ☐ Notice of Inventor's Patent Application
6) ☐ Other: _____

Applicant's amendments and comments, received September 25, 2008, have been fully considered by the examiner now of record. The following is a complete response to the September 22, 2008 communication.

Drawings

Figures 11-14 should be designated by a legend such as --Prior Art-- because only that which is old is illustrated. See MPEP § 608.02(g). Corrected drawings in compliance with 37 CFR 1.121(d) are required in reply to the Office action to avoid abandonment of the application. The replacement sheet(s) should be labeled "Replacement Sheet" in the page header (as per 37 CFR 1.84(c)) so as not to obstruct any portion of the drawing figures. If the changes are not accepted by the examiner, the applicant will be notified and informed of any required corrective action in the next Office action. The objection to the drawings will not be held in abeyance.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 9, 10, 23 and 24 are rejected under 35 U.S.C. 102(b) as being anticipated by Messing (2003/0004506).

Messing discloses an RF ablation catheter comprising a tip electrode (500 - Figure 5) including a unit which detects temperature (shown in Figure 5 without a label, but labeled as element 430 in Figure 4A) and a catheter shaft attached to the electrode (shown in Figure 1). The electrode is formed as a single metallic body and has a surface comprising three spherical or approximately spherical surfaces that have centers on the same straight line (Figure 5). A reduced diameter portion (520) is provided between each of the adjacent spherical surfaces such that the spherical surfaces are connected by a curved surface. Messing disclose the use of various materials for making the electrode, including the use of platinum, platinum-iridium alloys and gold (Para. 0042).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 11-22 are rejected under 35 U.S.C. 103(a) as being unpatentable over Messing ("506).

The Messing electrode has been addressed above. Messing fails to disclose the specific lengths and relative diameters for the electrode features. It is noted, however, that Messing teach that the reduced diameter portions provided between the electrodes serves to provide an electrode capable of creating a larger tissue lesion as disclosed by

applicant. See paragraph [0052] for discussion of the use of different diameter portions along the length of the electrode. The examiner maintains that one of ordinary skill in the art would obviously be capable, without undue experimentation, or arriving at desired lengths and diameters for the electrode depending on the particular procedure and the desired result. Applicant's specification fails to disclose any criticality or unexpected results associated with the particular parameters, and the examiner maintains that the skilled artisan would obviously be capable of optimizing the parameters to achieve a desired result.

To have provided the Messing electrode with any desired length and/or spacing for the spherical shaped elements would have been an obvious design consideration for one of ordinary skill in the art, particularly since the skilled artisan would recognize the obvious effects of modifying these parameters and the modification's impact on tissue treatments.

Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. He discloses another catheter having an electrode having a plurality of larger diameter portions (Figures 8A and 8B). Jackman et al (5,522,873) and Edwards et al (5,275,152) disclose electrodes that have two spherical elements joined in succession in a single electrode.

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP

§ 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael Peffley whose telephone number is (571) 272-4770. The examiner can normally be reached on Mon-Fri from 7am-4pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda Dvorak can be reached on (571) 272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Michael Peffley/
Primary Examiner, Art Unit 3739

/mp/
February 26, 2009